



中国认可
国际互认
检测
TESTING
CNAS L0599

Test Report

SL52045300659401TX

Date: November 04, 2020

Page 1 of 6

BENQ MEDICAL TECHNOLOGY CORP. TAICHUNG FACTORY
NO. 63 3F AND NO. 63, JINGKE CENTRAL RD., NANTUN DIST., TAICHUNG CITY 40852, TAIWAN (R.O.C.)

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)DELLCON FFP Respirator(Non-Sterile)

SGS Internal Ref.No. : HLA0345/2020

Sample Color : (A)WHITE

Style No. : DPPE-99

Previous No. : SL52035298560201TX

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Oct 20, 2020

Testing Period : Oct 20, 2020 - Nov 04, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, for further details, please refer to the following page(s).

Based on the results of the tests reported and the previous report SL52035298560201TX issued on Oct 09, 2020, our Lab concluded that the recommendation level can be as below.

Sample No.	Recommendation Level
(A)	FFP2 NR

Signed for and on behalf of
SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

3rd Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgs.com.cn
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com